

Certificate of Calibration

Report Number: 11-0803-05
Calibration Date: Wednesday, August 03, 2011

Customer: MicroDAQ.com, Ltd
Customer PO #: N/A
Sales Order #: N/A

Manufacturer: Mfg. Name Inserted Here
Model: Logger Model Number Inserted Here
Serial Number: Logger Serial Number Inserted Here

The above instrumentation has been calibrated and tested to meet or exceed the published specifications. This calibration and testing was performed using instrumentation and standards that are traceable to the United States Department of Commerce National Institute of Standards and Technology (NIST).

Calibration Environment

Temperature 22.3 °C
Humidity (RH) 29.0% RH

Calibration Information

| As Found Data: | Ref. Value | As Read | Diff. | Spec. |
|----------------|------------|-----------|----------|-----------|
| Pressure | 100.0 PSI | 100.0 PSI | +0.0 PSI | ± 3.0 PSI |

| As Left Data: | Condition Received | Condition Returned |
|---------------|--------------------|--------------------|
| Pressure | In Tolerance | In Tolerance |

Standards Used for Calibration

- EDGETECH Model DS2 Dew Point Hygrometer, Serial Number 33326, Sensor Serial Number 2921CRX. Calibration due 11/22/11
- EDGETECH Model DS2 Dew Point Hygrometer, Serial Number 36992, Sensor Serial Number 2A3593X. Calibration due 11/09/11
- KAYE Model M2801/IRTD-400, IPTS-68, Serial Number I0102. Calibration due 1/31/12
- AMETEK Pneumatic Pressure Tester, Model HK-500, Serial Number 82557. Calibration due 6/07/12
- FLUKE Voltage/Current Calibrator, Model 5500A Serial Number 9015013. Calibration due 12/07/11

Calibrated By: Peter Polizos
Title: Calibration & Product Technician

This Certificate/Report shall not be reproduced, except in full, without the written consent of MicroDAQ.com, Ltd.